

COUNTY OF FRESNO
ADDENDUM NUMBER: ONE (1)
RFQ NUMBER: 760-5333
GYRATORY COMPACTOR

February 27, 2015

PURCHASING USE
hrs

G:\PUBLIC\RFQ\FY 2014-15\760-5333 GYRATORY
COMPACTOR\760-5333 ADD 1.DOC

IMPORTANT: SUBMIT QUOTATION IN SEALED PACKAGE WITH QUOTATION NUMBER, CLOSING DATE AND BUYER'S NAME
MARKED CLEARLY ON THE OUTSIDE TO:

COUNTY OF FRESNO, Purchasing
4525 EAST HAMILTON AVENUE, 2ND FLOOR
FRESNO, CA 93702-4599

CLOSING DATE OF BID WILL BE AT 2:00 P.M., ON MARCH 17, 2015.

QUOTES WILL BE CONSIDERED LATE WHEN THE OFFICIAL PURCHASING TIME CLOCK READS 2:00 P.M.

All quotation information will be available for review after contract award.

Clarification of specifications is to be directed to: **Debbie Scharnick**, phone (559) 600-7111, e-mail
CountyPurchasing@co.fresno.ca.us, fax (559) 600-7126.

**NOTE THE ATTACHED ADDITIONS, DELETIONS AND/OR CHANGES TO THE REQUIREMENTS OF REQUEST
FOR QUOTATION NUMBER: 760-5333 AND INCLUDE THEM IN YOUR RESPONSE. PLEASE SIGN AND
RETURN THIS ADDENDUM WITH YOUR QUOTATION.**

ACKNOWLEDGMENT OF ADDENDUM NUMBER ONE (1) TO RFQ 760-5333

COMPANY NAME: _____
(PRINT)

SIGNATURE: _____

NAME & TITLE: _____
(PRINT)

Q1. The mold top is a part of our machine. Do you still want an additional 12 molds?

A1. If the mold top is part of the machine then I don't need 12 mold tops, but I do still need 12 molds.

Q2. Caltrans is also using 4" molds rather than 100 mm is this something you would also like to include in your quote?

A2. 4" molds may be substituted for 100 mm molds as long as the diameter of each is 3.995 to 4.000 inches. The internal angle verification kit needs to be compatible with these molds. The mold funnel and specimen extractor intended for use with the 100mm molds needs to work with the 4" molds instead. A standard for the 4" molds must be provided with the Bore gauge to verify the mold internal diameter.

Q3. The machine you have specified does not comply with what Caltrans is currently using. Their machine has shear measuring capabilities, do you want to include that in your quote?

A3. They can provide the shear adaptation as an alternative to distinguish them from other bidders and mention the compliance with similar Caltrans equipment.